



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***EQ-LAB, S.A. de C.V.***  
***Los Amoles #137, Hacienda San Miguel***  
***Guadalupe, Nuevo León, México. C.P. 67113***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

***Dimensional, Mass, Force and Weighing Devices, Mechanical, Thermodynamic, Chemical, Acoustic, Time and Frequency, Optical and Electrical Calibration***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

October 31, 2023

*Issue Date:*

October 31, 2023

*Expiration Date:*

January 31, 2026

*Accreditation No.:*

122472

*Certificate No.:*

L23-804

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
Guadalupe, Nuevo León, México. C.P. 67113

Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Micrometer Head <sup>F</sup>	Up to 50 mm	$(0.55 + 2.7 \times 10^{-3}L) \mu\text{m}$	Gauge Blocks Set	NMX-CH-093
Outside Micrometer <sup>F</sup>	Up to 2 000 mm	$(0.57 + 8.2 \times 10^{-3}L) \mu\text{m}$		
Inside Micrometer <sup>F</sup>	Up to 2 000 mm	$(0.55 + 2.7 \times 10^{-3}L) \mu\text{m}$		
Internal Micrometers with Three point (Holtest) <sup>F</sup>	6.3 mm to 100 mm	$(3.8 + 4.3 \times 10^{-2}L) \mu\text{m}$	Steel Setting Rings	NMX CH-92
Tubular Inside Micrometer Micrometer Head <sup>F</sup>	Up to 25 mm	$(0.57 + 1.0 \times 10^{-3}L) \mu\text{m}$	Gauge Blocks Set	JIS B 7508
Tubular Inside Micrometer Extensions Rods <sup>F</sup>	25 mm to 100 mm	$(0.92 + 2.8 \times 10^{-3}L) \mu\text{m}$		
Depth Micrometers <sup>F</sup>	Up to 300 mm	$(0.57 + 6.4 \times 10^{-3}L) \mu\text{m}$	Gauge Blocks Set	NMX-CH-093
Calipers, Vernier Dial and Digital <sup>F</sup>	Up to 2 000 mm	$(5.8 + 7.1 \times 10^{-3}L) \mu\text{m}$		
Linear Scales (Digital Rulers) <sup>FO</sup>	Up to 2 000 mm	$(5.8 + 7.1 \times 10^{-3}L) \mu\text{m}$		
Laser Distance Meter <sup>FO</sup>	127 mm to 2 006.6 mm	$(0.058 + 1.0 \times 10^{-6}L) \mu\text{m}$	Gague Blocks Set	ISO 16331-1
Telescope Gages (Only max and min values) <sup>FO</sup>	25 mm to 100 mm	$(0.73 + 3.7 \times 10^{-3}L) \mu\text{m}$	Gauge Blocks Set	NMX-CH-141
Height Gages Vernier, Dial and Digital <sup>F</sup>	Up to 1 000 mm	$(6.0 + 6.0 \times 10^{-3}L) \mu\text{m}$		
Height Master <sup>F</sup>	5 mm to 1 010 mm	$(1.6 + 6.6 \times 10^{-3}L) \mu\text{m}$	Gauge Blocks Set	NMX-CH-7863
Dial and Digital Indicators <sup>F</sup>	Up to 25 mm	$(0.99 + 6.9 \times 10^{-2}L) \mu\text{m}$	Dial Gage Tester	NMX-CH-463 NMX-CH-149
Dial Test Indicator <sup>F</sup>	Up to 1.6 mm	$(0.99 + 6.9 \times 10^{-2}L) \mu\text{m}$		
Dial and Digital Bore Gage <sup>F</sup>	1 mm to 25 mm	$(0.76 + 0.23L) \mu\text{m}$		
Pin and Plug Gages <sup>F</sup>	0.25 mm to 4.7 mm	$(0.79 + 8.0 \times 10^{-4}L) \mu\text{m}$	Digital Indicator	Euramet_cg-6
Standard Wire Gage <sup>F</sup>	0.25 mm to 4.7 mm	0.79 $\mu\text{m}$		
Setting Micrometer Standard <sup>F</sup>	25 mm to 100 mm	$(0.73 + 3.7 \times 10^{-3}L) \mu\text{m}$	Gague Blocks Set	JIS B 7545
Steel Thickness Gages <sup>F</sup>	0.04 mm to 1 mm	0.79 $\mu\text{m}$	Digital Indicator	JIS B 7524
Plastic Standard Coating Thickness Gages <sup>F</sup>	0.01 mm to 12.19 mm	$(0.79 + 8.0 \times 10^{-4}L) \mu\text{m}$	Digital Indicator	ASTM-D-1005



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
 Guadalupe, Nuevo León, México. C.P. 67113  
 Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Thread Plug Gages and Spline Gages Pitch Diameter <sup>F</sup>	2.1 mm to 46.7 mm	$(1.1 + 1.4 \times 10^{-2}L) \mu\text{m}$	Micrometer and Wire Set	ANSI B92.1 DIN 5480 ANSI/ASME B1.2
Coating Thickness Gages <sup>F</sup>	0.01 mm to 12.19 mm	5.9 $\mu\text{m}$	Foil Thickness	Standard ASTM D7091
Measure Tape <sup>F</sup>	Up to 50 m	$(0.029 + 6 \times 10^{-3}L) \text{mm}$	Rule Standard and Reticule	JIS B 7512
Pi Tape <sup>F</sup>	Up to 8 m	$(0.25 + 2 \times 10^{-5}L) \text{mm}$		
Rulers <sup>F</sup>	Up to 1 000 mm	$(0.025 + 2 \times 10^{-5}L) \text{mm}$		
Scantling Gages <sup>F</sup>	Up to 1 000 mm	0.29 mm	Rule Standard	
Surface Roughness Gage <sup>F</sup> (Ra)	1 $\mu\text{m}$ to 9.5 $\mu\text{m}$	$(0.77 + 1.2 \times 10^{-3}L) \mu\text{m}$	Precision Roughness Specimen	NMX-CH-12179
Surface Roughness Gage <sup>F</sup> (Rz)	6.6 $\mu\text{m}$ to 40 $\mu\text{m}$	$(0.79 + 1.8 \times 10^{-3}L) \mu\text{m}$		
Surface Roughness Gage <sup>F</sup> (Rt)	8 $\mu\text{m}$ to 43 $\mu\text{m}$	$(0.77 + 2.3 \times 10^{-3}L) \mu\text{m}$		
Surface Roughness Gage <sup>F</sup> (Rp)	1.9 $\mu\text{m}$ to 18 $\mu\text{m}$	$(0.78 + 3.7 \times 10^{-3}L) \mu\text{m}$		
Blocks Gages Grade 1 and 2 <sup>F</sup>	1 mm to 100 mm	$(0.63 + 3.7 \times 10^{-3}L) \mu\text{m}$	Blocks Standard Grade 0 Digital Indicator	NMX-CH-3650
4 Step Block Gage	6.35 mm to 25.4 mm	$(0.73 + 3.7 \times 10^{-3}L) \mu\text{m}$	Blocks Standard Grade 0 Digital Indicator	
Thickness Gages	Up to 25.4 mm	2.9 $\mu\text{m}$		
Protractor Goniometer Digital Level <sup>F</sup>	0° to 90°	$(0.049 + 1 \times 10^{-5}L)^\circ$	Angle Blocks Standard	NMX-CH-151
Squares <sup>F</sup>	0° to 90°	0.09°	Digital Protractor	ASME Y14.5
Squares Perpendicularity <sup>F</sup>	Up to 300 mm	0.011 mm	Test Indicator	NMX-CH-062
Profile Projector Angle error <sup>O</sup>	0° to 90°	$(0.049 + 1 \times 10^{-5}L)^\circ$	Angle Blocks Set	JIS B 7184
Profile Projectors X Axis error Y Axis error <sup>O</sup>	Up to 300 mm	$(1.1 + 0.15L) \mu\text{m}$	Standard Glass Scales	
Profile Projector - Magnification <sup>O</sup>	120 mm to 170 mm	$(0.055 + 4 \times 10^{-5}L) \text{mm}$		
Microscope and Vision System X Axis Linearity Y Axis Linearity <sup>O</sup>	Up to 300 mm	$(1.5 + 0.15L) \mu\text{m}$		



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
Guadalupe, Nuevo León, México. C.P. 67113

Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Surface Plates: Local Area Flatness <sup>O</sup>	250 mm x 250 mm to 2 500 mm x 1 600 mm	0.71 $\mu$ m	Repeat Reading Gage	JIS B 7513
Ultrasonic Thickness Gage <sup>F</sup>	6.35 to 25.4 mm	6.9 $\mu$ m	Step Block	ASTM E797
Levels <sup>F</sup>	Up to 300 mm Sensitivity 0.02 mm/m	0.046 mm/m	Reference Flat Surface	DIN 877

### Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Scales and Balances <sup>O</sup>	Up to 50 g (Res.= 0.05 mg)	0.064 mg	Weight Set OIML E2	NOM-010-SCFI
	Up to 100 g (Res.= 0.05 mg)	0.089 mg		
	Up to 200 g (Res.= 0.1 mg)	0.17 mg		
	Up to 500 g (Res.= 0.2 mg)	0.39 mg		
	Up to 1 kg (Res.= 2 mg)	3 mg	Weight Set OIML F1	NOM-010-SCFI
	Up to 2 kg (Res.= 5 mg)	6.4 mg		
	Up to 5 kg (Res.= 10 mg)	15 mg		
	Up to 10 kg (Res.= 2 g)	1.6 g	Weight Set OIML M1	NOM-010-SCFI
	Up to 20 kg (Res.= 5 g)	4 g		
	Up to 50 kg (Res.= 10 g)	8.1 g		
	Up to 100 kg (Res.= 20 g)	16 g		
	Up to 200 kg (Res.= 50 g)	40 g		
	Up to 500 kg (Res.= 100 g)	80 g		
	Up to 1 000 kg (Res.= 500 g)	0.4 kg		



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
Guadalupe, Nuevo León, México. C.P. 67113

Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Floor, Hopper & Scales	Up to 2 000 kg (Res.= 100 g)	0.8 kg	Weight Set OIML M1 and Material Substitution	CENAM Technical Guide
	Up to 6 000 kg (Res.= 1 kg)	2.4 kg		
	Up to 10 000 kg (Res.= 1 kg)	4 kg		
Weight OIML Class F1 and F2 and ASTM Class 3 and 4 <sup>F</sup>	1 g	0.033 mg	Weight Set OIML E2 Double Substitution	
	2 g	0.04 mg		
	5 g	0.05 mg		
	10 g	0.067 mg		
	20 g	0.083 mg		
	50 g	0.1 mg		
	100 g	0.17 mg		
	200 g	0.33 mg		
	500 g	0.83 mg		
	1 000 g	1.7 mg		
Weight OIML Class M1, M2 and M3, ASTM Class 5, 6 and 7 <sup>F</sup>	1 g	0.33 mg	Weight Set OIML E2 and F1 Double Substitution	
	2 g	0.4 mg		
	5 g	0.82 mg		
	10 g	0.83 mg		
	20 g	0.83 mg		
	50 g	0.84 mg		
	100 g	0.88 mg		
	200 g	0.99 mg		
	500 g	1.6 mg		
	1 kg	2.9 mg		
	2 kg	9.8 mg		
	5 kg	83 mg		
	10 kg	86 mg		
	20 kg	98 mg		
Force, Gages and Instruments Only Compression <sup>FO</sup>	100 N to 1 000 N	$(0.69 + 3.8 \times 10^{-3}F)$ % of reading	Load Cells	ISO-7500-1 ASTM E4
	1 000 N to 10 000 N	$(0.69 + 4 \times 10^{-4}F)$ % of reading		



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
Guadalupe, Nuevo León, México. C.P. 67113

Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Pressure / Vacuum Gages Air Medium and Water Medium <sup>FO</sup>	-14.5 psi to 0 psi	$(0.59 + 4 \times 10^{-4}P)$ psi	Pressure Gauge and Pressure Module Fluke	CENAM Technical Guide
	0 psi to 36 psi	0.059 psi		
	-1 psi to 1 psi	0.029 psi		
	Up to 30 psi	0.29 psi		
	Up to 300 psi	1.5 psi		
	Up to 1 000 psi	5.8 psi		
	Up to 10 000 psi	58 psi		
Dynamic Viscosity Meters @25°C <sup>FO</sup>	500 cP	2.1 cP	Cannon Standard Oil @ 25°C	ASTM D7042
	5 000 cP	24 cP		
	30 000 cP	162 cP		
Kinematic Viscosity Cups: Zahn, Ford, ISO, ASTM, DIN, Gradco, Shell, Frikmar Saybolt @25°C <sup>FO</sup>	36 cSt	0.1 cSt	Cannon Standard Oil	ASTM D1200 ASTM D4212
	126 cSt	0.29 cSt		
	1 000 cSt	1.3 cSt		
Indirect Verification of Rockwell Hardness Tester HRB <sup>O</sup>	20 HRB to 50 HRB	1.1 HRB	Test Block	ISO 6508-2 ASTM E18
	50 HRB to 80 HRB	1 HRB		
	80 HRB to 100 HRB	0.55 HRB		
Indirect Verification of Rockwell Hardness Tester HRC <sup>O</sup>	20 HRC to 30 HRC	0.48 HRC	Test Block	ASTM A956
	30 HRC to 55 HRC	0.57 HRC		
	55 HRC to 70 HRC	0.37 HRC		
Indirect Verification of Leeb Hardness Tester LD <sup>FO</sup>	365 HLD	16 HLD	Test Block	ASTM A956
Direct Verification of Durometer Shore Tester Types A, B, E & O <sup>F</sup> Durometer Force Spring Type M Type OO, OOO Type OOO-S Type C, D & DO	0.55 N to 8.05 N	0.000 14 N	Load Cell	ASTM D2240 ISO 21509
	4.445 N to 44.45 N	0.000 14 N		
	0.324 N to 0.765 N	0.000 14 N		
	0.203 N to 1.111 N	0.000 14 N		
	0.167 N to 1.932 N	0.000 14 N		
Micropipettes <sup>F</sup>	10 $\mu$ L to 200 $\mu$ L	0.38 $\mu$ L	Gravimetric Method Balances OHAUS	CENAM Technical Guide
	200 $\mu$ L to 500 $\mu$ L	0.39 $\mu$ L		





# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
 Guadalupe, Nuevo León, México. C.P. 67113  
 Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Micropipettes <sup>F</sup>	500 $\mu$ L to 1 000 $\mu$ L	0.42 $\mu$ L	Gravimetric Method Balances OHAUS	CENAM Technical Guide
	1 000 $\mu$ L to 2 000 $\mu$ L	0.52 $\mu$ L		
Pipettes <sup>F</sup>	10 mL to 200 mL	0.018 % of reading		
Burettes <sup>F</sup>	10 mL to 100 mL	0.018 % of reading		
Cylinders and Cubic Cups <sup>F</sup>	200 mL to 1 000 mL	0.02 % of reading		
Volumetric Flasks <sup>F</sup>	10 mL to 100 mL	0.018 % of reading		
	200 mL to 1 000 mL	$(0.02 + 2 \times 10^{-4}V)$ mL		
	1 000 mL to 2 000 mL	0.02 % of reading		
Beakers <sup>F</sup>	10 mL to 200 mL	0.018 % of reading		
	200 mL to 1 000 mL	$(0.02 + 2 \times 10^{-4}V)$ mL		
	1 000 mL to 2 000 mL	0.02 % of reading		
Pycnometers <sup>F</sup>	10 mL to 200 mL	0.018 % of reading		
Containers <sup>F</sup>	1 000 mL to 20 000 mL	0.02 % of reading		
Anemometer <sup>F</sup> (Air Flow)	Up to 9 m/s	$(0.58 + 0.23 L)$ m/s	Anemometer Comparison	IEC 61400-12-1 ASTM D5096
Torque Tools <sup>F</sup>	3.4 Nm to 17 Nm	$(0.066 + 2.6 \times 10^{-3} L)$ Nm	Torque Analyzer	Euramet_cg-14
	68 Nm to 339 Nm	$(0.75 + 6 \times 10^{-4} L)$ Nm		

### Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Infrared Guns, Pyrometers and Cameras <sup>F</sup>	50 °C to 1 000 °C	$(0.68 + 2.7 \times 10^{-3}T)$ °C	Precision Infrared Calibrator	CENAM Technical Guide
Bi-Metallic Thermometers <sup>F</sup>	-10 °C to 120 °C	0.83 °C	Dry-Well and Digital Thermometer w/ Pt-1000	Euramet cg-20
	30 °C to 600 °C	$(0.79 + 2.8 \times 10^{-3}T)$ °C	Fluke 725 w/TC-k	
Climatic Chambers: Ovens, Furnaces, Mufflers, Incubators, Refrigerators, Freezers, Cold Rooms Error of Indication <sup>O</sup>	0 °C to 60 °C	$(0.2 + 2.4 \times 10^{-3}T)$ °C	Reference Thermo Hygrometer	Euramet cg-20



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
 Guadalupe, Nuevo León, México. C.P. 67113  
 Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Thermodynamic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Climatic Chambers: Ovens, Furnaces, Mufflers, Incubators, Refrigerators, Freezers, Cold Rooms Error of Indication <sup>O</sup>	-40 °C to 200 °C	$(0.58 + 4.0 \times 10^{-5}T) \text{ } ^\circ\text{C}$	Digital Thermometer w/Pt-1000	Euramet cg-20
	8 °C to 800 °C	$(1.1 + 2.0 \times 10^{-3}T) \text{ } ^\circ\text{C}$	Fluke 725 w/TC-k	
RTDs, Rods and Probes with Thermocouple J, K, T, E, R, S, B, L U and N <sup>F</sup>	-10 °C to 120 °C	0.83 °C	Dry-Well and Digital Thermometer w/ Pt-1000	
	35 °C to 600 °C	$(0.57 + 3.0 \times 10^{-3}T) \text{ } ^\circ\text{C}$	Dry-Well and Fluke 725 w/TC-k	
Liquid in Glass Thermometers <sup>F</sup>	-25 °C to 150 °C	$(0.14 + 1.0 \times 10^{-4}T) \text{ } ^\circ\text{C}$	Bath Circulator and Digital Thermometer w/ Pt-1000	
Thermo Hygrometer <sup>F</sup> (Humidity)	11.5 % RH to 97.5 % RH	$(1.5 + 2.3 \times 10^{-3}H) \text{ } \% \text{ RH}$	Reference Thermohygrometer and Humidity Chamber	
Climatic Chambers (Humidity) Sensors and Recorders <sup>FO</sup>	11.5 % RH to 97.5 % RH	$(0.97 + 2.3 \times 10^{-3}H) \text{ } \% \text{ RH}$		
Thermo Hygrometer (Temperature) <sup>F</sup>	2 °C to 60 °C	0.31 °C	Reference Thermohygrometer and Temperature Chamber	
Heating Plates <sup>FO</sup>	Up to 260 °C	$(1.1 + 4 \times 10^{-4}T) \text{ } ^\circ\text{C}$	Digital Thermometer Fluke 54 w/TC-K	
Dry Well <sup>F</sup>	-40 °C to 200 °C	$(0.58 + 4 \times 10^{-5}T) \text{ } ^\circ\text{C}$	Digital Thermometer w/Pt-1000	
Baths Circulators <sup>FO</sup>	-40 °C to 200 °C	$(0.58 + 5 \times 10^{-5}T) \text{ } ^\circ\text{C}$	Digital Thermometer w/Pt-1000	
	35 °C to 600 °C	$(0.55 + 3.1 \times 10^{-3}T) \text{ } ^\circ\text{C}$	Dry-Well and Fluke 725 w/TC-k	

### Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
pH Meter <sup>FO</sup>	4 pH	0.062 pH	Buffer Solutions	NMX-CH-166 CEM QU-003
	7 pH	0.061 pH		
	10 pH	0.082 pH		





## Certificate of Accreditation: Supplement

### EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
Guadalupe, Nuevo León, México. C.P. 67113

Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

#### Chemical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Conductivity Meter <sup>FO</sup>	1 415 $\mu$ S/cm	15 $\mu$ S/cm	Buffer Solution	OIML R 68
Karl Fisher Titration Equipment <sup>O</sup>	1 mg/g	0.013 mg/g	Standard Solutions	ASTM E 203-01
	10 mg/g	0.19 mg/g		

#### Acoustic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Acoustic Level Generate (F=1 kHz) <sup>FO</sup>	94 dB	0.37 dB	Acoustic Calibrator	IEC 61672-1 NMX-CH-389-1-IMNC
	104 dB	0.4 dB		
	114 dB	0.43 dB		
Acoustic Level Measurement (F=1 kHz) <sup>FO</sup>	94 dB	0.79 dB	Acoustic Calibrator Sonometer of Reference	IEC 61672-1 NMX-CH-389-1-IMNC
	114 dB	0.87 dB		

#### Time and Frequency

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment to Output Frequency <sup>FO</sup>	113 rpm to 28 800 rpm	$(0.53 + 6 \times 10^{-4}V)$ rpm	LUTRON DT-2259 Digital Tachometer / Stroboscope AS432B	CENAM Technical Guide
Equipment to Output Time <sup>FO</sup>	1 s to 36 000 s	$(0.49 + 4 \times 10^{-6}t)$ s	JUNSD Stopwatch ITTC-7.6-02-07	

#### Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Ev Light Meters <sup>F</sup>	120 lux to 3 000 lux	$(1.1 + 2.4 \times 10^{-2} Ev)$ lux	Luxometer	CENAM Technical Guide
Ev Illuminance <sup>F</sup>	260 lux to 1 880 lux	$(1 + 2.5 \times 10^{-2} Ev)$ lux	Luxometer	ASTM E1164
Refractive Index <sup>F</sup>	1 °Brix to 85 °Brix	$(0.51 + 2 \times 10^{-4}B)$ % Brix	Sucrose Standards	OIML R-108
Gloss/Specular Reflectance Meter Angle of Incline <sup>FO</sup>	20° / 93.9 Gloss Units 60° / 96.5 Gloss Units 85° / 99.7 Gloss Units	0.2 Gloss Units 0.2 Gloss Units 0.23 Gloss Units	Ceram Research Gloss Standard	ASTM D523-14



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
 Guadalupe, Nuevo León, México. C.P. 67113  
 Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	
Spectrophotometers Transmittance <sup>FO</sup>	$\tau$ : 10 %	$(0.056 + 2 \times 10^{-5}\tau) \% \tau$	Neutral Density Filters, Holmium Oxide Glass	CENAM Technical Guide	
	$\tau$ : 20 %	$(0.061 + 4 \times 10^{-5}\tau) \% \tau$			
	$\tau$ : 30 %	$(0.067 + 6 \times 10^{-5}\tau) \% \tau$			
	$\lambda$ : 440 nm to 750 nm	0.05 nm			
$\rho(\lambda)$ Spectral Reflectance <sup>FO</sup>	Color Values:		Ceramic Research Tiles	CENAM Technical Guide ASTM E-1164	
	CIE L:	9 to 93 Units			$(1.8 + 1.6 \times 10^{-2}L)$ Units
	CIE a*:	-27 to 29 Units			$(0.45 + 3.6 \times 10^{-3}a^*)$ Units
	CIE b*:	-15 to 55 Units			$(0.5 + 4.9 \times 10^{-3}b^*)$ Units

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
High Voltage DC Output <sup>O</sup>	2.5 kV to 35 kV	0.42 kV	Fluke 80K-40 w / Fluke 289	CENAM Technical Guide
High Voltage AC Output (60 Hz) <sup>O</sup>	2.5 kV to 25 kV	0.31 kV		
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type B <sup>FO</sup>	600 °C to 1 800 °C	0.46 °C	Fluke 725 Electrical Simulation of Thermocouple Output	Euramet_cg-11
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type E <sup>FO</sup>	-200 °C to 950 °C	0.18 °C		
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type J <sup>FO</sup>	-200 °C to 1 200 °C	0.2 °C		
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type K <sup>FO</sup>	-200 °C to 1 370 °C	0.22 °C		
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type L <sup>FO</sup>	-200 °C to 900 °C	0.2 °C		



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
 Guadalupe, Nuevo León, México. C.P. 67113  
 Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type N <sup>FO</sup>	-200 °C to 400 °C	0.11 °C	Fluke 725 Electrical Simulation of Thermocouple Output	Euramet_cg-11
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type R <sup>FO</sup>	-20 °C to 1 750 °C	0.48 °C		
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type S <sup>FO</sup>	-20 °C to 1 750 °C	0.46 °C		
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type T <sup>FO</sup>	-250 °C to 400 °C	0.18 °C		
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type U <sup>FO</sup>	-200 °C to 600 °C	0.11 °C		
Temperature Calibration, Indication, and Control Equipment used with RTD Type Pt 385, 100 $\Omega$ <sup>FO</sup>	-200 °C to 800 °C	0.11 °C	Fluke 725 Electrical Simulation of RTD Output	Euramet_cg-11
Temperature Calibration, Indication, and Control Equipment used with RTD Type Pt Ni 385, 120 $\Omega$ (Ni 120) <sup>FO</sup>	-80°C to 260°C	0.18 °C		
Temperature Calibration, Indication, and Control Equipment used with RTD Type Pt 385, 200 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.11 °C		
Temperature Calibration, Indication, and Control Equipment used with RTD Type Pt 385, 500 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.11 °C		
Temperature Calibration, Indication, and Control Equipment used with RTD Type Pt 385, 1 000 $\Omega$ <sup>FO</sup>	-200 °C to 630 °C	0.11 °C		



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
 Guadalupe, Nuevo León, México. C.P. 67113  
 Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Temperature Calibration, Indication, and Control Equipment used with RTD Type Pt 3 916, 100 $\Omega^{\text{FO}}$	-200 °C to 630 °C	0.11 °C	Fluke 725 Electrical Simulation of RTD Output	Euramet_cg-11
	Temperature Calibration, Indication, and Control Equipment used with RTD Type Pt 3 926, 100 $\Omega^{\text{FO}}$	-200 °C to 630 °C		
Equipment Output DC Voltage <sup>FO</sup>	10 mV to 100 mV	0.012 mV	Fluke 725	CENAM Technical Guide
	3 V to 30 V	0.006 2 V	Fluke 289	
	5 mV to 50 mV	0.004 mV		
	50 mV to 500 mV	0.17 mV		
	1 V to 5 V	0.000 27 V		
	5 V to 50 V	0.001 7 V		
Equipment Output AC Voltage 60 Hz <sup>FO</sup>	50 V to 500 V	1.4 V	Fluke 289	
	100 V to 1000 V	5.8 V		
Equipment Output DC Current	4 mA to 24 mA	0.019 mA	Fluke 725	
	50 $\mu$ A to 500 $\mu$ A	0.55 $\mu$ A	Fluke 289	
	500 $\mu$ A to 5000 $\mu$ A	40 $\mu$ A		
	5 mA to 50 mA	0.005 5 mA		
	40 mA to 400 mA	0.19 mA		
Equipment Output AC Current 60 HzV	1 A to 5 A	0.000 34 A	Fluke 289	
	1 A to 10 A	0.001 3 A		
	1 mA to 10 mA	0.005 9 mA		
	10 mA to 100 mA	0.013 mA		
	100 mA to 1000 mA	0.11 mA		
	1 A to 10 A	0.005 9 A		
	10 A to 100 A	0.013 A		
	100 A to 1000 A	0.11 A		
	1000 A to 2000 A	0.26 A		
Equipment Output Resistance <sup>FO</sup>	15 $\Omega$ to 3 200 $\Omega$	4.1 $\Omega$	Fluke 725	
	5 $\Omega$ to 50 $\Omega$	0.036 $\Omega$	Fluke 289	
	50 $\Omega$ to 500 $\Omega$	0.64 $\Omega$		



# Certificate of Accreditation: Supplement

## EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
 Guadalupe, Nuevo León, México. C.P. 67113  
 Contact Name Roberto Delgado Andrade Phone: 813-067-7022

Accreditation is granted to the facility to perform the following calibrations:

### Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY ( $\pm$ )	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Equipment Output Resistance <sup>FO</sup>	1 k $\Omega$ to 5 k $\Omega$	0.000 068 k $\Omega$	Fluke 289	CENAM Technical Guide
	5 k $\Omega$ to 50 k $\Omega$	0.004 k $\Omega$		
	50 k $\Omega$ to 500 k $\Omega$	0.41 k $\Omega$		
	1 M $\Omega$ to 5 M $\Omega$	0.027 M $\Omega$		
	5 M $\Omega$ to 50 M $\Omega$	2.8 M $\Omega$		
	50 M $\Omega$ to 500 M $\Omega$	75 M $\Omega$		
Equipment Output Frequency <sup>FO</sup>	50 Hz to 900 Hz	0.008 5 Hz	Fluke 725	
	1 kHz to 10 kHz	0.001 2 kHz		
Equipment Input DC Current <sup>FO</sup>	4 mA to 24 mA	0.000 58 mA		
Equipment Input Resistance <sup>FO</sup>	15 $\Omega$ to 3200 $\Omega$	4.1 $\Omega$		
Equipment Input Frequency <sup>FO</sup>	50 Hz to 900 Hz	0.008 2 Hz		
	1 kHz to 10 kHz	0.000 13 kHz		

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor  $k$  (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript O means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer<sup>O</sup> would mean that the laboratory performs this calibration onsite at the customer's location.



## Certificate of Accreditation: Supplement

### EQ-LAB, S.A. de C.V.

Los Amoles #137, Hacienda San Miguel  
Guadalupe, Nuevo León, México. C.P. 67113

Contact Name Roberto Delgado Andrade Phone: 813-067-7022

*Accreditation is granted to the facility to perform the following calibrations:*

5. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer<sup>FO</sup> would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
6. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
7. The term L represents length in inches or millimeters as appropriate to the uncertainty statement.
8. The term T represents temperature in °C or °F as appropriate to the uncertainty statement.
9. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.
10. The term H represent Relative Humidity in units of RH as appropriate to the uncertainty statement.
11. The term V represents Volume in liters or milliliters (including SI multiple and submultiple units) as appropriate to the uncertainty statement.
12. The term V represents velocity in rpm
13. The term t represents time in seconds (s) or millisecond (ms) as appropriate to the uncertainty statement.
14. The term P represents pressure in psi.
15. The term F represents force applied in Newton or kilogram-force